



Retraction notice

Retraction notice to “Surface dose rate variations in planar and curved geometries of $^{106}\text{Ru}/^{106}\text{Rh}$ plaque sources for ocular tumors” [Phys. Med. 89 (2021) 200–209]

A.K. Bakshi^{a,c,*}, Vandana Shrivastava^a, Arghya Chattaraj^a, Kasthuri Samuel^a,
T. Palani Selvam^{a,c}, B.K. Sapa^{a,c}, Prithwish Sinharoy^{b,c}, Dayamoy Banerjee^{b,c}, G. Sugilala^{b,c},
Smitha Manohar^{a,b,c}, C.P. Kaushik^{b,c}

^a Radiological Physics & Advisory Division, India

^b Nuclear Recycle Group, Bhabha Atomic Research Centre, Mumbai 400 085, India

^c Homi Bhabha National Institute, Mumbai 400 094, India

This article has been retracted: please see Elsevier Policy on Article Withdrawal (<http://www.elsevier.com/locate/withdrawalpolicy>).

This article has been retracted at the request of the Editor-in-Chief and Authors for the following reasons:

1. In EBT3 film-based measurements of dose profiles at different depths involving a curved source, stack of films were placed below the source and both the source and the films were in contact geometry (Ref Figure 5b). However, in these measurements, air present in the curved

region of the source was not filled with the film to represent clinical application. It is felt that the % dose distributions could be different when air region is filled with film as beta particles will be attenuated significantly.

2. Similarly, the Monte Carlo-based depth dose calculations utilised similar geometry, i.e. air was present in the curved region of the source (Ref Figures 7 & 8).

DOI of original article: <https://doi.org/10.1016/j.ejmp.2021.08.001>.

* Corresponding authors at: Radiological Physics & Advisory Division, India.

E-mail address: akbakshi@barc.gov.in (A.K. Bakshi).

<https://doi.org/10.1016/j.ejmp.2022.02.016>

Available online 23 February 2022

1120-1797/© 2022 Associazione Italiana di Fisica Medica e Sanitaria. Published by Elsevier Ltd. All rights reserved.

